

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121623\
 Data File : VU056939.D
 Acq On : 16 Dec 2023 20:05
 Operator : MD/SY
 Sample : 05810-16
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F7C91

Quant Time: Dec 16 23:49:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Dec 15 01:53:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	388812	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	401096	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	172559	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	66803	2.004	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	40.000%
7) Chloroethane-d5	1.908	69	58021	2.326	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	46.600%#
11) 1,1-Dichloroethene-d2	2.560	65	29901	2.097	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	42.000%#
20) 2-Butanone-d5	4.621	46	198768	30.200	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	60.400%
24) Chloroform-d	5.055	84	163068	2.584	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	51.600%#
26) 1,2-Dichloroethane-d4	5.695	65	80575	2.698	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	54.000%#
32) Benzene-d6	5.721	84	292906	2.354	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	47.000%#
36) 1,2-Dichloropropane-d6	6.682	67	99370	2.552	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	51.000%#
41) Toluene-d8	7.894	98	230947	2.077	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	41.600%#
43) trans-1,3-Dichloroprop...	8.177	79	34698	2.298	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	46.000%#
46) 2-Hexanone-d5	8.631	63	146894	25.792	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	51.580%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	72296	2.566	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	51.400%#
66) 1,2-Dichlorobenzene-d4	12.190	152	88943	2.622	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	52.400%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121623\
Data File : VU056939.D
Acq On : 16 Dec 2023 20:05
Operator : MD/SY
Sample : 05810-16
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F7C91

Quant Time: Dec 16 23:49:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Dec 15 01:53:53 2023
Response via : Initial Calibration

