

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121623\
 Data File : VU056951.D
 Acq On : 17 Dec 2023 03:04
 Operator : MD/SY
 Sample : 05811-10
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F7B06

Quant Time: Dec 17 03:19:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Dec 15 01:53:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	195046	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	189582	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	90336	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	67341	4.028	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	80.600%
7) Chloroethane-d5	1.907	69	59036	4.719	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.400%
11) 1,1-Dichloroethene-d2	2.560	65	31174	4.358	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	87.200%
20) 2-Butanone-d5	4.621	46	202704	61.394	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	122.780%
24) Chloroform-d	5.055	84	166863	5.271	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.400%
26) 1,2-Dichloroethane-d4	5.695	65	81990	5.473	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	109.400%
32) Benzene-d6	5.721	84	299093	5.086	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.800%
36) 1,2-Dichloropropane-d6	6.685	67	102001	5.542	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	110.800%
41) Toluene-d8	7.894	98	230585	4.387	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.800%
43) trans-1,3-Dichloroprop...	8.177	79	34039	4.769	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.400%
46) 2-Hexanone-d5	8.630	63	154154	57.264	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	114.520%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	79114	5.941	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	118.800%
66) 1,2-Dichlorobenzene-d4	12.187	152	91464	5.151	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.000%
Target Compounds						Qvalue
3) Chloromethane	1.515	50	6123	0.305	ug/L	98
47) Tetrachloroethene	8.550	164	2519	0.197	ug/L	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121623\
Data File : VU056951.D
Acq On : 17 Dec 2023 03:04
Operator : MD/SY
Sample : 05811-10
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F7B06

Quant Time: Dec 17 03:19:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Dec 15 01:53:53 2023
Response via : Initial Calibration

