

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121623\
 Data File : VU056953.D
 Acq On : 17 Dec 2023 03:53
 Operator : MD/SY
 Sample : 05811-12
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F7B13

Quant Time: Dec 17 04:25:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Dec 15 01:53:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	260997	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	267278	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	103885	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	62404	2.789	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	55.800%	
7) Chloroethane-d5	1.907	69	57494	3.434	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	68.600%	
11) 1,1-Dichloroethene-d2	2.563	65	27045	2.825	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	56.600%#	
20) 2-Butanone-d5	4.621	46	193146	43.717	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	87.440%	
24) Chloroform-d	5.055	84	156529	3.695	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	73.800%	
26) 1,2-Dichloroethane-d4	5.695	65	78307	3.906	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	78.200%	
32) Benzene-d6	5.721	84	279424	3.371	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	67.400%#	
36) 1,2-Dichloropropane-d6	6.682	67	96149	3.705	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	74.200%	
41) Toluene-d8	7.894	98	206806	2.791	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	55.800%#	
43) trans-1,3-Dichloroprop...	8.177	79	31347	3.115	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	62.400%	
46) 2-Hexanone-d5	8.631	63	142279	37.489	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	74.980%	
56) 1,1,2,2-Tetrachloroeth...	10.749	84	74172	3.951	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	79.000%	
66) 1,2-Dichlorobenzene-d4	12.187	152	71886	3.520	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	70.400%#	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121623\
Data File : VU056953.D
Acq On : 17 Dec 2023 03:53
Operator : MD/SY
Sample : 05811-12
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F7B13

Quant Time: Dec 17 04:25:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Dec 15 01:53:53 2023
Response via : Initial Calibration

