

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121718\
 Data File : VU028709.D
 Acq On : 17 Dec 2018 10:17
 Operator : MD/SY
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 18 03:21:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U120618W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 07 00:39:42 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	77	0.00
2 T	Dichlorodifluoromethane	0.670	0.663	1.0	76	0.00
3 P	Chloromethane	1.232	1.000	18.8	65	0.00
4 C	Vinyl Chloride	0.889	0.851	4.3#	74	0.00
5 T	Bromomethane	0.358	0.311	13.1	72	0.00
6 T	Chloroethane	0.503	0.500	0.6	75	0.00
7 T	Trichlorofluoromethane	0.936	0.939	-0.3	79	0.00
8 T	Diethyl Ether	0.438	0.400	8.7	73	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.560	0.556	0.7	78	0.00
10 T	Methyl Iodide	0.439	0.472	-7.5	81	0.00
11 T	Tert butyl alcohol	0.170	0.149	12.4	68	0.00
12 CM	1,1-Dichloroethene	0.557	0.519	6.8#	73	0.00
13 T	Acrolein	0.136	0.123	9.6	69	0.00
14 T	Allyl chloride	1.317	1.215	7.7	75	0.00
15 T	Acrylonitrile	0.472	0.461	2.3	74	0.00
16 T	Acetone	0.498	0.472	5.2	78	0.00
17 T	Carbon Disulfide	2.005	1.889	5.8	75	0.00
18 T	Methyl Acetate	1.291	1.321	-2.3	78	0.00
19 T	Methyl tert-butyl Ether	2.110	2.066	2.1	76	0.00
20 T	Methylene Chloride	0.734	0.671	8.6	76	0.00
21 T	trans-1,2-Dichloroethene	0.627	0.594	5.3	74	0.00
22 T	Diisopropyl ether	2.561	2.563	-0.1	76	0.00
23 T	Vinyl Acetate	2.112	2.075	1.8	74	0.00
24 P	1,1-Dichloroethane	1.341	1.305	2.7	77	0.00
25 T	2-Butanone	0.680	0.660	2.9	74	0.00
26 T	2,2-Dichloropropane	1.062	1.036	2.4	78	0.00
27 T	cis-1,2-Dichloroethene	0.709	0.708	0.1	78	0.00
28 T	Bromochloromethane	0.609	0.648	-6.4	82	0.00
29 T	Tetrahydrofuran	0.437	0.426	2.5	73	0.00
30 C	Chloroform	1.168	1.181	-1.1#	79	0.00
31 T	Cyclohexane	1.573	1.269	19.3	75	0.00
32 T	1,1,1-Trichloroethane	0.970	0.961	0.9	77	0.00
33 S	1,2-Dichloroethane-d4	0.763	0.772	-1.2	81	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	74	0.00
35 S	Dibromofluoromethane	0.315	0.347	-10.2	83	0.00
36 T	1,1-Dichloropropene	0.547	0.562	-2.7	76	0.00
37 T	Ethyl Acetate	0.707	0.716	-1.3	73	0.00
38 T	Carbon Tetrachloride	0.441	0.481	-9.1	82	0.00
39 T	Methylcyclohexane	0.668	0.662	0.9	75	0.00
40 TM	Benzene	1.611	1.679	-4.2	77	0.00
41 T	Methacrylonitrile	0.387	0.412	-6.5	76	0.00
42 TM	1,2-Dichloroethane	0.564	0.591	-4.8	78	0.00
43 T	Isopropyl Acetate	1.119	1.149	-2.7	74	0.00
44 TM	Trichloroethene	0.358	0.373	-4.2	78	0.00
45 C	1,2-Dichloropropane	0.453	0.482	-6.4#	76	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121718\
 Data File : VU028709.D
 Acq On : 17 Dec 2018 10:17
 Operator : MD/SY
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 18 03:21:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U120618W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 07 00:39:42 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.269	0.281	-4.5	76	0.00
47 T	Bromodichloromethane	0.523	0.568	-8.6	80	0.00
48 T	Methyl methacrylate	0.556	0.579	-4.1	74	0.00
49 T	1,4-Dioxane	0.009	0.009	0.0	78	0.00
50 S	Toluene-d8	1.245	1.341	-7.7	81	0.00
51 T	4-Methyl-2-Pentanone	0.706	0.732	-3.7	75	0.00
52 CM	Toluene	0.948	0.975	-2.8#	75	0.00
53 T	t-1,3-Dichloropropene	0.623	0.666	-6.9	78	0.00
54 T	cis-1,3-Dichloropropene	0.676	0.724	-7.1	78	0.00
55 T	1,1,2-Trichloroethane	0.375	0.400	-6.7	76	0.00
56 T	Ethyl methacrylate	0.642	0.670	-4.4	74	0.00
57 T	1,3-Dichloropropane	0.696	0.727	-4.5	77	0.00
58 T	2-Chloroethyl Vinyl ether	0.329	0.353	-7.3	78	0.00
59 T	2-Hexanone	0.559	0.571	-2.1	75	0.00
60 T	Dibromochloromethane	0.367	0.412	-12.3	84	0.00
61 T	1,2-Dibromoethane	0.393	0.422	-7.4	77	0.00
62 S	4-Bromofluorobenzene	0.465	0.497	-6.9	82	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	74	0.00
64 T	Tetrachloroethene	0.334	0.339	-1.5	77	0.00
65 PM	Chlorobenzene	1.065	1.110	-4.2	78	0.00
66 T	1,1,1,2-Tetrachloroethane	0.340	0.389	-14.4	83	0.00
67 C	Ethyl Benzene	1.959	2.018	-3.0#	77	0.00
68 T	m/p-Xylenes	0.713	0.740	-3.8	75	0.00
69 T	o-Xylene	0.693	0.726	-4.8	77	0.00
70 T	Styrene	1.146	1.206	-5.2	77	0.00
71 P	Bromoform	0.289	0.356	-23.2#	89	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	74	0.00
73 T	Isopropylbenzene	3.837	3.897	-1.6	77	0.00
74 T	N-amyl acetate	2.213	2.158	2.5	73	0.00
75 P	1,1,2,2-Tetrachloroethane	1.474	1.544	-4.7	78	0.00
76 T	1,2,3-Trichloropropane	1.277	1.341	-5.0	77	0.00
77 T	Bromobenzene	0.897	0.931	-3.8	78	0.00
78 T	n-propylbenzene	4.767	4.930	-3.4	78	0.00
79 T	2-Chlorotoluene	2.746	2.863	-4.3	78	0.00
80 T	1,3,5-Trimethylbenzene	3.176	3.315	-4.4	77	0.00
81 T	trans-1,4-Dichloro-2-butene	0.550	0.585	-6.4	91	0.00
82 T	4-Chlorotoluene	3.254	3.326	-2.2	78	0.00
83 T	tert-Butylbenzene	3.061	3.217	-5.1	79	0.00
84 T	1,2,4-Trimethylbenzene	3.273	3.431	-4.8	78	0.00
85 T	sec-Butylbenzene	3.933	4.047	-2.9	76	0.00
86 T	p-Isopropyltoluene	3.341	3.430	-2.7	77	0.00
87 T	1,3-Dichlorobenzene	1.694	1.742	-2.8	79	0.00
88 T	1,4-Dichlorobenzene	1.753	1.799	-2.6	80	0.00
89 T	n-Butylbenzene	3.381	3.467	-2.5	76	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121718\
Data File : VU028709.D
Acq On : 17 Dec 2018 10:17
Operator : MD/SY
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_U
LabSampleId :
VSTDCCC050

Quant Time: Dec 18 03:21:39 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U120618W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 07 00:39:42 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.505	0.627	-24.2#	93	0.00
91 T	1,2-Dichlorobenzene	1.695	1.738	-2.5	79	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.319	0.343	-7.5	78	0.00
93 T	1,2,4-Trichlorobenzene	1.215	1.187	2.3	78	0.00
94 T	Hexachlorobutadiene	0.555	0.550	0.9	77	0.00
95 T	Naphthalene	4.067	3.988	1.9	75	0.00
96 T	1,2,3-Trichlorobenzene	1.222	1.224	-0.2	79	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6