

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121723\
 Data File : VU056965.D
 Acq On : 17 Dec 2023 13:57
 Operator : MD/SY
 Sample : VU1217WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK107

Quant Time: Dec 18 01:16:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Dec 18 01:16:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	271534	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	284425	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	109896	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	102614	4.409	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.200%
7) Chloroethane-d5	1.907	69	68257	3.919	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	78.400%
11) 1,1-Dichloroethene-d2	2.560	65	46649	4.684	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	93.600%
20) 2-Butanone-d5	4.624	46	179199	38.986	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	77.980%
24) Chloroform-d	5.055	84	188468	4.276	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.600%
26) 1,2-Dichloroethane-d4	5.695	65	87029	4.173	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.400%
32) Benzene-d6	5.721	84	370699	4.202	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.000%
36) 1,2-Dichloropropane-d6	6.685	67	113546	4.112	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	82.200%
41) Toluene-d8	7.894	98	303338	3.847	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	77.000%
43) trans-1,3-Dichloroprop...	8.177	79	38975	3.640	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	72.800%
46) 2-Hexanone-d5	8.631	63	139923	34.645	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	69.300%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	77415	3.875	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	77.400%
66) 1,2-Dichlorobenzene-d4	12.190	152	93571	4.331	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	86.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121723\
Data File : VU056965.D
Acq On : 17 Dec 2023 13:57
Operator : MD/SY
Sample : VU1217WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK107

Quant Time: Dec 18 01:16:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Dec 18 01:16:15 2023
Response via : Initial Calibration

