

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121723\
 Data File : VU056978.D
 Acq On : 17 Dec 2023 19:18
 Operator : MD/SY
 Sample : 05842-13
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F7C26

Quant Time: Dec 18 01:19:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Dec 18 01:16:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	237711	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	261167	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	104917	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	62942	3.089	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	61.800%
7) Chloroethane-d5	1.904	69	62948	4.128	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	82.600%
11) 1,1-Dichloroethene-d2	2.560	65	29693	3.406	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.200%
20) 2-Butanone-d5	4.615	46	211045	52.448	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.900%
24) Chloroform-d	5.055	84	167869	4.351	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.000%
26) 1,2-Dichloroethane-d4	5.695	65	84535	4.630	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.600%
32) Benzene-d6	5.721	84	294906	3.641	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	72.800%
36) 1,2-Dichloropropane-d6	6.682	67	99387	3.920	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	78.400%
41) Toluene-d8	7.894	98	223519	3.087	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	61.800%#
43) trans-1,3-Dichloroprop...	8.174	79	31836	3.238	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	64.800%
46) 2-Hexanone-d5	8.627	63	155963	42.056	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.120%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	80566	4.392	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	87.800%
66) 1,2-Dichlorobenzene-d4	12.190	152	83127	4.031	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	80.600%
Target Compounds					Qvalue	
13) Acetone	2.621	43	75956	30.974	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121723\
Data File : VU056978.D
Acq On : 17 Dec 2023 19:18
Operator : MD/SY
Sample : 05842-13
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F7C26

Quant Time: Dec 18 01:19:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Dec 18 01:16:15 2023
Response via : Initial Calibration

