

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121723\
 Data File : VU056980.D
 Acq On : 17 Dec 2023 20:06
 Operator : MD/SY
 Sample : 05842-15
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F7C63

Quant Time: Dec 18 01:20:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Dec 18 01:16:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	259295	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	255496	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	104819	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	66860	3.008	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	60.200%	
7) Chloroethane-d5	1.907	69	65360	3.930	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	78.600%	
11) 1,1-Dichloroethene-d2	2.560	65	29737	3.127	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	62.600%	
20) 2-Butanone-d5	4.615	46	212219	48.349	ug/L	-0.01
Spiked Amount	50.000	Range 40 - 130	Recovery	=	96.700%	
24) Chloroform-d	5.055	84	170297	4.046	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	81.000%	
26) 1,2-Dichloroethane-d4	5.695	65	85840	4.310	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	86.200%	
32) Benzene-d6	5.721	84	296791	3.745	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	75.000%	
36) 1,2-Dichloropropane-d6	6.685	67	103389	4.168	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	83.400%	
41) Toluene-d8	7.894	98	206124	2.910	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	58.200%#	
43) trans-1,3-Dichloroprop...	8.177	79	30307	3.151	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	63.000%	
46) 2-Hexanone-d5	8.627	63	153106	42.202	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	84.400%	
56) 1,1,2,2-Tetrachloroeth...	10.749	84	77954	4.344	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	86.800%	
66) 1,2-Dichlorobenzene-d4	12.190	152	87281	4.236	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	84.800%	
Target Compounds						
					Qvalue	
13) Acetone	2.624	43	77466	28.960	ug/L	100
21) 2-Butanone	4.737	43	7419	1.645	ug/L #	53

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121723\
Data File : VU056980.D
Acq On : 17 Dec 2023 20:06
Operator : MD/SY
Sample : 05842-15
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F7C63

Quant Time: Dec 18 01:20:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Dec 18 01:16:15 2023
Response via : Initial Calibration

