

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121724\
 Data File : VU062384.D
 Acq On : 17 Dec 2024 13:17
 Operator : MD/SY
 Sample : P5228-05
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Dec 18 00:14:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Dec 18 00:11:18 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	94999	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	90497	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	42222	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	27997	4.369	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	87.400%
7) Chloroethane-d5	1.907	69	22331	4.900	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	98.000%
11) 1,1-Dichloroethene-d2	2.557	65	13978	4.585	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	91.800%
20) 2-Butanone-d5	4.631	46	55759	45.405	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	90.800%
24) Chloroform-d	5.052	84	65582	4.729	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.600%
26) 1,2-Dichloroethane-d4	5.695	65	36664	5.376	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.600%
32) Benzene-d6	5.717	84	123093	4.723	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.400%
36) 1,2-Dichloropropane-d6	6.682	67	35551	4.767	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.400%
41) Toluene-d8	7.891	98	112260	4.559	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.200%
43) trans-1,3-Dichloroprop...	8.177	79	14976	4.562	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.200%
46) 2-Hexanone-d5	8.627	63	43757	44.184	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.360%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	27651	4.765	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.200%
66) 1,2-Dichlorobenzene-d4	12.187	152	41750	5.366	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121724\
Data File : VU062384.D
Acq On : 17 Dec 2024 13:17
Operator : MD/SY
Sample : P5228-05
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Dec 18 00:14:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Dec 18 00:11:18 2024
Response via : Initial Calibration

