

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121724\
 Data File : VU062403.D
 Acq On : 17 Dec 2024 22:35
 Operator : MD/SY
 Sample : P5310-17
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0CP8

Quant Time: Dec 18 01:14:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Dec 18 01:12:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	97467	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	91821	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	45090	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	27301	4.153	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	83.000%
7) Chloroethane-d5	1.908	69	21602	4.620	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.400%
11) 1,1-Dichloroethene-d2	2.560	65	13737	4.392	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	87.800%
20) 2-Butanone-d5	4.624	46	65306	51.833	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.660%
24) Chloroform-d	5.052	84	67315	4.731	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.600%
26) 1,2-Dichloroethane-d4	5.692	65	37135	5.308	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.200%
32) Benzene-d6	5.718	84	117500	4.443	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.800%
36) 1,2-Dichloropropane-d6	6.682	67	34191	4.519	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.400%
41) Toluene-d8	7.891	98	108207	4.331	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.600%
43) trans-1,3-Dichloroprop...	8.174	79	14835	4.454	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.000%
46) 2-Hexanone-d5	8.628	63	49405	49.168	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.340%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	29879	5.075	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.400%
66) 1,2-Dichlorobenzene-d4	12.187	152	42473	5.112	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121724\
Data File : VU062403.D
Acq On : 17 Dec 2024 22:35
Operator : MD/SY
Sample : P5310-17
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0CP8

Quant Time: Dec 18 01:14:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Dec 18 01:12:26 2024
Response via : Initial Calibration

