

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121724\
 Data File : VU062381.D
 Acq On : 17 Dec 2024 12:04
 Operator : MD/SY
 Sample : P5274-08
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Dec 18 00:13:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Dec 18 00:11:18 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	97330	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	93447	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	43700	5.000	ug/L	0.00

System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	27150	4.135	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.800%
7) Chloroethane-d5	1.907	69	21575	4.620	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.400%
11) 1,1-Dichloroethene-d2	2.560	65	13954	4.468	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	89.400%
20) 2-Butanone-d5	4.628	46	55172	43.851	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	87.700%
24) Chloroform-d	5.055	84	64726	4.555	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.200%
26) 1,2-Dichloroethane-d4	5.692	65	33778	4.835	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.600%
32) Benzene-d6	5.717	84	120395	4.473	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.400%
36) 1,2-Dichloropropane-d6	6.682	67	34082	4.426	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.600%
41) Toluene-d8	7.891	98	109549	4.309	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.200%
43) trans-1,3-Dichloroprop...	8.174	79	14394	4.246	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.000%
46) 2-Hexanone-d5	8.627	63	42360	41.423	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	82.840%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	25644	4.280	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	85.600%
66) 1,2-Dichlorobenzene-d4	12.183	152	38374	4.765	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121724\
Data File : VU062381.D
Acq On : 17 Dec 2024 12:04
Operator : MD/SY
Sample : P5274-08
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Dec 18 00:13:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Dec 18 00:11:18 2024
Response via : Initial Calibration

