

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121724\
 Data File : VU062409.D
 Acq On : 18 Dec 2024 01:03
 Operator : MD/SY
 Sample : P5309-03
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0CF2

Quant Time: Dec 18 01:34:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Dec 18 01:12:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	97157	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	92484	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	44709	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	24673	3.765	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	75.200%
7) Chloroethane-d5	1.907	69	20812	4.465	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.200%
11) 1,1-Dichloroethene-d2	2.557	65	13071	4.193	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	83.800%
20) 2-Butanone-d5	4.628	46	61629	49.070	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.140%
24) Chloroform-d	5.052	84	62980	4.441	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.800%
26) 1,2-Dichloroethane-d4	5.692	65	33875	4.857	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.200%
32) Benzene-d6	5.717	84	113110	4.246	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.000%
36) 1,2-Dichloropropane-d6	6.682	67	33520	4.398	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.000%
41) Toluene-d8	7.891	98	105656	4.199	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.000%
43) trans-1,3-Dichloroprop...	8.174	79	13787	4.110	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.200%
46) 2-Hexanone-d5	8.627	63	46116	45.566	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	91.140%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	28727	4.844	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.800%
66) 1,2-Dichlorobenzene-d4	12.187	152	40286	4.890	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.800%
Target Compounds						
13) Acetone	2.650	43	4349	5.086	ug/L	97
22) cis-1,2-Dichloroethene	4.660	96	7306	0.967	ug/L	86


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121724\
Data File : VU062409.D
Acq On : 18 Dec 2024 01:03
Operator : MD/SY
Sample : P5309-03
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0CF2

Quant Time: Dec 18 01:34:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Dec 18 01:12:26 2024
Response via : Initial Calibration





#13

Acetone

Concen: 5.086 ug/L

RT: 2.650 min Scan# 42

Delta R.T. 0.010 min

Lab File: VU062409.D

Acq: 18 Dec 2024 01:03

Instrument :

MSVOA_U

ClientSampleId :





C0CF2

Tgt Ion: 43 Resp: 4349

Ion Ratio Lower Upper

43 100

58 30.6 0.0 64.0



#22

cis-1,2-Dichloroethene

Concen: 0.967 ug/L

RT: 4.660 min Scan# 1048

Delta R.T. 0.006 min

Lab File: VU062409.D

Acq: 18 Dec 2024 01:03

Tgt Ion: 96 Resp: 7306

Ion Ratio Lower Upper

96 100

61 110.7 92.0 170.8

98 59.4 44.5 82.7

