

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U121819\
 Data File : VU036268.D
 Acq On : 18 Dec 2019 13:10
 Operator : JC/MD
 Sample : K6275-16DL 5X
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFDT2DL

Quant Time: Dec 19 01:12:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121319WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Dec 19 00:50:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	269534	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	290891	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.84	152	93575	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.62	65	123454	5.20	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	104.00%
7) Chloroethane-d5	1.93	69	110742	5.21	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.20%
11) 1,1-Dichloroethene-d2	2.59	63	127512	3.72	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.40%
20) 2-Butanone-d5	4.68	46	188368	42.36	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	84.72%
24) Chloroform-d	5.11	84	172839	4.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.60%
26) 1,2-Dichloroethane-d4	5.75	65	89977	4.49	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.80%
32) Benzene-d6	5.77	84	324065	4.43	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.60%
36) 1,2-Dichloropropane-d6	6.73	67	105590	4.58	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.60%
41) Toluene-d8	7.93	98	266383	3.87	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	77.40%
43) trans-1,3-Dichloropropene-	8.21	79	37402	3.95	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.00%
46) 2-Hexanone-d5	8.68	63	140620	38.79	ug/L	0.01
Spiked Amount	50.000	Range	45 - 130	Recovery	=	77.58%
57) 1,1,2,2-Tetrachloroethane-	10.78	84	90321	4.29	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	85.80%
64) 1,2-Dichlorobenzene-d4	12.22	152	83212	4.64	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.80%

Target Compounds

					Ovalue
13) Acetone	2.67	43	8833	2.279 ug/L	95
16) Methylene chloride	3.09	84	15966	0.543 ug/L	92
22) cis-1,2-Dichloroethene	4.72	96	31694	1.411 ug/L	99
34) Trichloroethene	6.58	95	185680	8.056 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU121819\
Data File : VU036268.D
Acq On : 18 Dec 2019 13:10
Operator : JC/MD
Sample : K6275-16DL 5X
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFDT2DL

Quant Time: Dec 19 01:12:04 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121319WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Dec 19 00:50:55 2019
Response via : Initial Calibration

