

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU121820\
 Data File : VU041684.D
 Acq On : 18 Dec 2020 17:24
 Operator : SY/MD
 Sample : L5142-06
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 ETZB5

Quant Time: Dec 19 08:57:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Dec 19 04:12:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	121342	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	109408	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	50785	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	35916	3.79	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	75.80%
7) Chloroethane-d5	1.92	69	34617	4.71	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.20%
11) 1,1-Dichloroethene-d2	2.58	63	54753	3.13	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.60%
20) 2-Butanone-d5	4.66	46	149696	54.93	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.86%
24) Chloroform-d	5.08	84	74874	4.61	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.20%
26) 1,2-Dichloroethane-d4	5.72	65	45165	4.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.00%
32) Benzene-d6	5.74	84	148940	4.75	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.00%
36) 1,2-Dichloropropane-d6	6.70	67	48987	4.99	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.80%
41) Toluene-d8	7.91	98	126372	4.47	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.40%
43) trans-1,3-Dichloropropene-	8.19	79	21002	4.76	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.20%
46) 2-Hexanone-d5	8.64	63	131948	54.65	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	109.30%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	48590	5.45	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	109.00%
64) 1,2-Dichlorobenzene-d4	12.19	152	46591	5.16	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.20%

Target Compounds					Ovalue
3) Chloromethane	1.53	50	867	0.087 ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU121820\
Data File : VU041684.D
Acq On : 18 Dec 2020 17:24
Operator : SY/MD
Sample : L5142-06
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ETZB5

Quant Time: Dec 19 08:57:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Dec 19 04:12:21 2020
Response via : Initial Calibration

