

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU121820\
 Data File : VU041682.D
 Acq On : 18 Dec 2020 16:36
 Operator : SY/MD
 Sample : VIBLK36
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK36

Quant Time: Dec 19 08:54:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Dec 19 04:12:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	104456	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	98128	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	42398	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	33493	4.10	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.00%
7) Chloroethane-d5	1.92	69	33915	5.37	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	107.40%
11) 1,1-Dichloroethene-d2	2.58	63	51031	3.39	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.80%
20) 2-Butanone-d5	4.66	46	137381	58.57	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	117.14%
24) Chloroform-d	5.08	84	70512	5.04	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.80%
26) 1,2-Dichloroethane-d4	5.72	65	46547	5.87	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	117.40%
32) Benzene-d6	5.74	84	144963	5.16	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.20%
36) 1,2-Dichloropropane-d6	6.70	67	44871	5.09	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.80%
41) Toluene-d8	7.90	98	114829	4.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.60%
43) trans-1,3-Dichloropropene-	8.18	79	16859	4.26	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.20%
46) 2-Hexanone-d5	8.64	63	119055	54.98	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	109.96%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	43962	5.49	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	109.80%
64) 1,2-Dichlorobenzene-d4	12.19	152	42015	5.57	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	111.40%


Target Compounds					Ovalue
16) Methylene chloride	3.06	84	3135	0.370 ug/L	87




(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU121820\
Data File : VU041682.D
Acq On : 18 Dec 2020 16:36
Operator : SY/MD
Sample : VIBLK36
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK36

Quant Time: Dec 19 08:54:59 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Dec 19 04:12:21 2020
Response via : Initial Calibration





#16
Methylene chloride
Concen: 0.370 ug/L
RT: 3.06 min Scan# 535
Delta R.T. 0.00 min
Lab File: VU041682.D
Acq: 18 Dec 2020 16:36

Instrument :
MSVOA_U
ClientSampleId :
VIBLK36

Tot Ion: 84 Resp: 3135
Ion Ratio Lower Upper
84 100
86 59.1 45.1 83.7
49 100.8 83.1 154.3

