

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121823\
 Data File : VU056986.D
 Acq On : 18 Dec 2023 12:07
 Operator : MD/SY
 Sample : VU1218WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK108

Quant Time: Dec 18 14:31:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Dec 18 01:16:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	229649	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	237726	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	91268	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	95690	4.861	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	97.200%
7) Chloroethane-d5	1.907	69	69651	4.728	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.600%
11) 1,1-Dichloroethene-d2	2.560	65	40049	4.755	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	95.200%
20) 2-Butanone-d5	4.621	46	168100	43.242	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	86.480%
24) Chloroform-d	5.055	84	171579	4.603	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.000%
26) 1,2-Dichloroethane-d4	5.695	65	83297	4.722	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.400%
32) Benzene-d6	5.721	84	327164	4.437	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.800%
36) 1,2-Dichloropropane-d6	6.682	67	102779	4.453	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.000%
41) Toluene-d8	7.894	98	272042	4.128	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.600%
43) trans-1,3-Dichloroprop...	8.174	79	35902	4.011	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.200%
46) 2-Hexanone-d5	8.631	63	132857	39.358	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	78.720%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	70888	4.245	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	85.000%
66) 1,2-Dichlorobenzene-d4	12.190	152	89370	4.981	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121823\
Data File : VU056986.D
Acq On : 18 Dec 2023 12:07
Operator : MD/SY
Sample : VU1218WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK108

Quant Time: Dec 18 14:31:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Dec 18 01:16:15 2023
Response via : Initial Calibration

