

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121823\
 Data File : VU057011.D
 Acq On : 18 Dec 2023 23:16
 Operator : MD/SY
 Sample : 05771-03DL 40X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E0YA0DL

Quant Time: Dec 19 05:11:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Dec 19 02:36:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	281818	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	297635	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	125807	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	95320	3.946	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	79.000%
7) Chloroethane-d5	1.907	69	70199	3.883	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	77.600%
11) 1,1-Dichloroethene-d2	2.557	65	44551	4.311	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	86.200%
20) 2-Butanone-d5	4.634	46	206761	43.341	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	86.680%
24) Chloroform-d	5.055	84	187465	4.098	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.000%
26) 1,2-Dichloroethane-d4	5.695	65	88285	4.079	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.600%
32) Benzene-d6	5.718	84	354324	3.838	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	76.800%
36) 1,2-Dichloropropane-d6	6.682	67	114131	3.950	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	79.000%
41) Toluene-d8	7.894	98	294809	3.573	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	71.400%
43) trans-1,3-Dichloroprop...	8.174	79	39705	3.543	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	70.800%
46) 2-Hexanone-d5	8.637	63	159086	37.642	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	75.280%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	87084	4.165	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.400%
66) 1,2-Dichlorobenzene-d4	12.190	152	103972	4.204	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	84.000%
Target Compounds						
16) Methylene chloride	3.042	84	9262	0.372	ug/L	Qvalue 96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121823\
Data File : VU057011.D
Acq On : 18 Dec 2023 23:16
Operator : MD/SY
Sample : 05771-03DL 40X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E0YA0DL

Quant Time: Dec 19 05:11:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Dec 19 02:36:03 2023
Response via : Initial Calibration





#16
Methylene chloride
Concen: 0.372 ug/L
RT: 3.042 min Scan# 54
Delta R.T. 0.003 min
Lab File: VU057011.D
Acq: 18 Dec 2023 23:16

Instrument :
MSVOA_U
ClientSampleId :
E0YA0DL

Tgt Ion: 84 Resp: 9262
Ion Ratio Lower Upper
84 100
86 72.0 43.6 81.0
49 115.0 80.6 149.6

