

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU121918\
 Data File : VU028786.D
 Acq On : 19 Dec 2018 12:07
 Operator : MD/SY
 Sample : J6449-14
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 FB-2018-12-13

Quant Time: Dec 20 05:56:06 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U120618W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 07 00:39:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	128378	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	218024	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	198087	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	78714	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	85267	43.52	ug/l	0.00
Spiked Amount			Recovery	=	87.04%	
35) Dibromofluoromethane	4.88	113	71369	52.02	ug/l	0.00
Spiked Amount			Recovery	=	104.04%	
50) Toluene-d8	7.56	98	274214	50.52	ug/l	0.00
Spiked Amount			Recovery	=	101.04%	
62) 4-Bromofluorobenzene	10.30	95	88615	43.68	ug/l	0.00
Spiked Amount			Recovery	=	87.36%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU121918\
Data File : VU028786.D
Acq On : 19 Dec 2018 12:07
Operator : MD/SY
Sample : J6449-14
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
FB-2018-12-13

Quant Time: Dec 20 05:56:06 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U120618W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 07 00:39:42 2018
Response via : Initial Calibration

