

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU121918\
 Data File : VU028788.D
 Acq On : 19 Dec 2018 12:55
 Operator : MD/SY
 Sample : J6405-02DL 5X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TT101D1-20181212DL

Quant Time: Dec 20 06:55:34 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U120618W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 07 00:39:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	125503	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.88	114	213170	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	194713	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	77435	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	82429	43.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.08%	
35) Dibromofluoromethane	4.88	113	69699	51.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.92%	
50) Toluene-d8	7.56	98	270222	50.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.84%	
62) 4-Bromofluorobenzene	10.30	95	85430	43.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.14%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.29	101	3571	2.543	ug/l	97
12) 1,1-Dichloroethene	2.28	96	1409	1.008	ug/l #	73
44) Trichloroethene	6.19	130	59274	38.802	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU121918\
Data File : VU028788.D
Acq On : 19 Dec 2018 12:55
Operator : MD/SY
Sample : J6405-02DL 5X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TT101D1-20181212DL

Quant Time: Dec 20 06:55:34 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U120618W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 07 00:39:42 2018
Response via : Initial Calibration

