

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU121918\
 Data File : VU028790.D
 Acq On : 19 Dec 2018 13:44
 Operator : MD/SY
 Sample : J6474-04
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 S49-0493

Quant Time: Dec 20 06:03:32 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U120618W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 07 00:39:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	122871	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.88	114	211856	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	190151	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	76626	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	81366	43.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.78%	
35) Dibromofluoromethane	4.88	113	69246	51.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.88%	
50) Toluene-d8	7.56	98	266016	50.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.88%	
62) 4-Bromofluorobenzene	10.30	95	84996	43.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.24%	
Target Compounds						
16) Acetone	2.32	43	3786	3.094	ug/l	96
27) cis-1,2-Dichloroethene	4.23	96	5583	3.202	ug/l	78
44) Trichloroethene	6.18	130	92451	60.896	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU121918\
Data File : VU028790.D
Acq On : 19 Dec 2018 13:44
Operator : MD/SY
Sample : J6474-04
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
S49-0493

Quant Time: Dec 20 06:03:32 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U120618W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 07 00:39:42 2018
Response via : Initial Calibration

