

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU121918\
 Data File : VU028801.D
 Acq On : 19 Dec 2018 18:09
 Operator : MD/SY
 Sample : J6449-09
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-17C_R1

Quant Time: Dec 20 06:53:25 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U120618W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 07 00:39:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	122621	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.88	114	207207	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	189530	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	74850	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	81450	43.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.06%	
35) Dibromofluoromethane	4.88	113	68450	52.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.98%	
50) Toluene-d8	7.56	98	262216	50.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.66%	
62) 4-Bromofluorobenzene	10.30	95	82760	42.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.86%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.23	96	2816	1.619	ug/l	88
64) Tetrachloroethene	8.23	164	677	0.536	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU121918\
Data File : VU028801.D
Acq On : 19 Dec 2018 18:09
Operator : MD/SY
Sample : J6449-09
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-17C_R1

Quant Time: Dec 20 06:53:25 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U120618W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 07 00:39:42 2018
Response via : Initial Calibration

