

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U121919\
 Data File : VU036296.D
 Acq On : 19 Dec 2019 20:56
 Operator : JC/MD
 Sample : VU1219WBL02
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK05

Quant Time: Dec 20 05:41:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121319WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Dec 20 05:33:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	264178	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	291328	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	110129	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.62	65	115676	4.97	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	99.40%
7) Chloroethane-d5	1.93	69	104085	4.99	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.80%
11) 1,1-Dichloroethene-d2	2.60	63	120437	3.59	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.80%
20) 2-Butanone-d5	4.68	46	188089	43.16	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	86.32%
24) Chloroform-d	5.11	84	163825	4.38	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.60%
26) 1,2-Dichloroethane-d4	5.75	65	90798	4.62	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.40%
32) Benzene-d6	5.77	84	320707	4.38	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.60%
36) 1,2-Dichloropropane-d6	6.73	67	102885	4.46	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.20%
41) Toluene-d8	7.93	98	261420	3.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.80%
43) trans-1,3-Dichloropropene-	8.21	79	37295	3.93	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.60%
46) 2-Hexanone-d5	8.67	63	151764	41.81	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.62%
57) 1,1,2,2-Tetrachloroethane-	10.78	84	90926	4.31	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.20%
64) 1,2-Dichlorobenzene-d4	12.22	152	100298	4.76	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.20%

Target Compounds				Ovalue
13) Acetone	2.66	43	9964	2.623 ug/L 99
16) Methylene chloride	3.08	84	7471	0.259 ug/L 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU121919\
Data File : VU036296.D
Acq On : 19 Dec 2019 20:56
Operator : JC/MD
Sample : VU1219WBL02
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK05

Quant Time: Dec 20 05:41:02 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121319WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Dec 20 05:33:55 2019
Response via : Initial Calibration

