

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121922\
 Data File : VU052452.D
 Acq On : 19 Dec 2022 10:54
 Operator : JC/MD
 Sample : VSTDCCC0.5
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD0.5096

Quant Time: Dec 20 00:27:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM120122.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Dec 02 00:27:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.207	114	16883	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.349	117	15102	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.745	152	7147	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.599	65	9507	0.478	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery =	96.000%		
4) 1,2-Dichloroethane-d4	5.661	65	7847	0.416	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	84.000%		
7) 1,2-Dichloropropane-d6	6.639	67	9132	0.503	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery =	100.000%		
8) Toluene-d8	7.833	98	11368	0.416	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	84.000%		
10) 1,1,2,2-Tetrachloroeth...	10.682	84	9742	0.502	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery =	100.000%		
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.606	62	9888	0.459	ug/L	88
6) Trichloroethene	6.489	95	8260	0.489	ug/L	99
9) 1,2-Dibromoethane	8.854	107	6310	0.497	ug/L	99
12) 1,2,3-Trichloropropane	10.751	75	5603	0.469	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.910	75	1088	0.455	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121922\
Data File : VU052452.D
Acq On : 19 Dec 2022 10:54
Operator : JC/MD
Sample : VSTDCCC0.5
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD0.5096

Quant Time: Dec 20 00:27:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM120122.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Dec 02 00:27:38 2022
Response via : Initial Calibration

