

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121923\
 Data File : VU057050.D
 Acq On : 19 Dec 2023 17:35
 Operator : MD/SY
 Sample : 05843-15
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F7C93

Quant Time: Dec 20 04:59:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Dec 20 04:54:16 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	208553	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	210886	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	92287	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	68491	3.831	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	76.600%
7) Chloroethane-d5	1.907	69	55727	4.166	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	83.400%
11) 1,1-Dichloroethene-d2	2.563	65	28996	3.791	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.800%
20) 2-Butanone-d5	4.621	46	128791	36.481	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	72.960%
24) Chloroform-d	5.055	84	141984	4.194	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.800%
26) 1,2-Dichloroethane-d4	5.695	65	68906	4.302	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.000%
32) Benzene-d6	5.721	84	253365	3.873	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	77.400%
36) 1,2-Dichloropropane-d6	6.682	67	78314	3.825	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	76.600%
41) Toluene-d8	7.894	98	213581	3.653	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	73.000%
43) trans-1,3-Dichloroprop...	8.177	79	29309	3.691	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	73.800%
46) 2-Hexanone-d5	8.631	63	102382	34.190	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	68.380%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	57114	3.856	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	77.200%
66) 1,2-Dichlorobenzene-d4	12.187	152	73958	4.077	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	81.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121923\
Data File : VU057050.D
Acq On : 19 Dec 2023 17:35
Operator : MD/SY
Sample : 05843-15
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F7C93

Quant Time: Dec 20 04:59:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Dec 20 04:54:16 2023
Response via : Initial Calibration

