

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121923\
 Data File : VU057062.D
 Acq On : 19 Dec 2023 22:49
 Operator : MD/SY
 Sample : 05865-05
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F7C11

Quant Time: Dec 20 06:55:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Dec 20 06:35:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	201651	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	203942	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	87252	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	68528	3.965	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	79.200%
7) Chloroethane-d5	1.910	69	54058	4.179	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	83.600%
11) 1,1-Dichloroethene-d2	2.563	65	29125	3.938	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.800%
20) 2-Butanone-d5	4.624	46	127811	37.443	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	74.880%
24) Chloroform-d	5.055	84	139139	4.251	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.000%
26) 1,2-Dichloroethane-d4	5.695	65	68701	4.436	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.800%
32) Benzene-d6	5.721	84	252410	3.990	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.800%
36) 1,2-Dichloropropane-d6	6.685	67	76542	3.866	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	77.400%
41) Toluene-d8	7.894	98	212955	3.767	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.400%
43) trans-1,3-Dichloroprop...	8.177	79	30005	3.908	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.200%
46) 2-Hexanone-d5	8.630	63	103568	35.764	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	71.520%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	56380	3.936	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	78.800%
66) 1,2-Dichlorobenzene-d4	12.190	152	72481	4.226	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	84.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121923\
Data File : VU057062.D
Acq On : 19 Dec 2023 22:49
Operator : MD/SY
Sample : 05865-05
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F7C11

Quant Time: Dec 20 06:55:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Dec 20 06:35:40 2023
Response via : Initial Calibration

