

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122023\
 Data File : VU057095.D
 Acq On : 20 Dec 2023 16:28
 Operator : MD/SY
 Sample : 05866-04
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F7C34

Quant Time: Dec 20 23:46:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Dec 20 23:41:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	196514	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	202521	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	87670	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	57660	3.423	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	68.400%
7) Chloroethane-d5	1.911	69	49157	3.900	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	78.000%
11) 1,1-Dichloroethene-d2	2.563	65	26660	3.699	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.000%
20) 2-Butanone-d5	4.624	46	125656	37.774	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	75.540%
24) Chloroform-d	5.055	84	134589	4.219	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.400%
26) 1,2-Dichloroethane-d4	5.695	65	65861	4.363	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.200%
32) Benzene-d6	5.721	84	230308	3.666	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	73.400%
36) 1,2-Dichloropropane-d6	6.685	67	72911	3.708	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	74.200%
41) Toluene-d8	7.894	98	198210	3.530	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	70.600%
43) trans-1,3-Dichloroprop...	8.174	79	27200	3.567	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	71.400%
46) 2-Hexanone-d5	8.631	63	98295	34.181	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	68.360%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	55964	3.934	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	78.600%
66) 1,2-Dichlorobenzene-d4	12.187	152	72068	4.182	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	83.600%
Target Compounds						Qvalue
17) Methyl tert-butyl Ether	3.358	73	8668	0.245	ug/L	# 92
47) Tetrachloroethene	8.547	164	4007	0.293	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122023\
Data File : VU057095.D
Acq On : 20 Dec 2023 16:28
Operator : MD/SY
Sample : 05866-04
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F7C34

Quant Time: Dec 20 23:46:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Dec 20 23:41:59 2023
Response via : Initial Calibration

