

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122023\
 Data File : VU057118.D
 Acq On : 21 Dec 2023 02:40
 Operator : MD/SY
 Sample : 05892-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F7B07

Quant Time: Dec 21 05:03:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Dec 21 01:34:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	176215	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	178443	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	81329	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	52344	3.465	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	69.400%
7) Chloroethane-d5	1.911	69	46447	4.109	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	82.200%
11) 1,1-Dichloroethene-d2	2.560	65	25707	3.978	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	79.600%
20) 2-Butanone-d5	4.647	46	72042	24.151	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	48.300%
24) Chloroform-d	5.059	84	123747	4.326	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.600%
26) 1,2-Dichloroethane-d4	5.695	65	57516	4.249	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.000%
32) Benzene-d6	5.721	84	212643	3.842	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	76.800%
36) 1,2-Dichloropropane-d6	6.685	67	68184	3.936	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	78.800%
41) Toluene-d8	7.894	98	186171	3.763	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.200%
43) trans-1,3-Dichloroprop...	8.177	79	23195	3.453	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	69.000%
46) 2-Hexanone-d5	8.631	63	85155	33.607	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	67.220%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	48886	3.900	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	78.000%
66) 1,2-Dichlorobenzene-d4	12.187	152	71091	4.447	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	89.000%
Target Compounds						Qvalue
14) Carbon disulfide	2.789	76	9111	0.184	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122023\
Data File : VU057118.D
Acq On : 21 Dec 2023 02:40
Operator : MD/SY
Sample : 05892-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F7B07

Quant Time: Dec 21 05:03:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Dec 21 01:34:13 2023
Response via : Initial Calibration

