

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122023\
 Data File : VU057093.D
 Acq On : 20 Dec 2023 15:40
 Operator : MD/SY
 Sample : 05866-02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F7C08

Quant Time: Dec 20 23:45:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Dec 20 23:41:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	187744	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	194190	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	83227	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	64847	4.030	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	80.600%
7) Chloroethane-d5	1.911	69	55389	4.599	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.000%
11) 1,1-Dichloroethene-d2	2.563	65	29349	4.263	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	85.200%
20) 2-Butanone-d5	4.624	46	134861	42.435	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	84.860%
24) Chloroform-d	5.055	84	142157	4.665	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.200%
26) 1,2-Dichloroethane-d4	5.695	65	70354	4.879	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.600%
32) Benzene-d6	5.721	84	251550	4.176	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.600%
36) 1,2-Dichloropropane-d6	6.682	67	78846	4.182	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	83.600%
41) Toluene-d8	7.891	98	211324	3.925	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.600%
43) trans-1,3-Dichloroprop...	8.177	79	29619	4.051	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.000%
46) 2-Hexanone-d5	8.631	63	104839	38.020	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	76.040%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	60731	4.452	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.000%
66) 1,2-Dichlorobenzene-d4	12.187	152	75826	4.635	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122023\
Data File : VU057093.D
Acq On : 20 Dec 2023 15:40
Operator : MD/SY
Sample : 05866-02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F7C08

Quant Time: Dec 20 23:45:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Dec 20 23:41:59 2023
Response via : Initial Calibration

