

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU122118\
 Data File : VU028892.D
 Acq On : 21 Dec 2018 22:41
 Operator : MD/SY
 Sample : J6509-07
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 ETO88

Quant Time: Dec 22 03:31:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM122018WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Dec 22 01:55:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	173379	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	163852	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	65728	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	61213	44.04	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	88.08%
7) Chloroethane-d5	1.67	69	50952	47.46	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	94.92%
11) 1,1-Dichloroethene-d2	2.27	63	83008	35.51	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.02%
21) 2-Butanone-d5	4.18	46	106917	109.79	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	109.79%
24) Chloroform-d	4.65	84	107702	49.28	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.56%
26) 1,2-Dichloroethane-d4	5.31	65	68826	51.08	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.16%
32) Benzene-d6	5.34	84	238229	51.46	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.92%
36) 1,2-Dichloropropane-d6	6.33	67	83091	52.84	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	105.68%
41) Toluene-d8	7.56	98	209241	49.88	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.76%
43) trans-1,3-Dichloropropene-	7.85	79	34206	48.54	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.08%
47) 2-Hexanone-d5	8.31	63	83053	111.20	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	111.20%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	110911	50.84	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	101.68%
64) 1,2-Dichlorobenzene-d4	11.86	152	72315	54.40	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.80%


Target Compounds					Ovalue
13) Acetone	2.32	43	2726	2.747 ug/L	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU122118\
Data File : VU028892.D
Acq On : 21 Dec 2018 22:41
Operator : MD/SY
Sample : J6509-07
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ETO88

Quant Time: Dec 22 03:31:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM122018WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Dec 22 01:55:05 2018
Response via : Initial Calibration





#13
 Acetone
 Concen: 2.747 ug/L
 RT: 2.32 min Scan# 357
 Delta R.T. 0.00 min
 Lab File: VU028892.D
 Acq: 21 Dec 2018 22:41

Instrument :
 MSVOA_U
 ClientSampleId :
 ETO88

Tot Ion: 43 Resp: 2726
 Ion Ratio Lower Upper
 43 100
 58 33.7 0.0 70.2

