

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U122118\
 Data File : VU028906.D
 Acq On : 22 Dec 2018 04:20
 Operator : MD/SY
 Sample : J6542-07
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F9F70

Quant Time: Dec 22 06:07:49 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM122018WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Dec 22 01:55:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	167408	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	156642	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	61623	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	57156	42.59	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	85.18%
7) Chloroethane-d5	1.67	69	46839	45.18	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	90.36%
11) 1,1-Dichloroethene-d2	2.27	63	75775	33.57	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.14%
21) 2-Butanone-d5	4.18	46	99263	105.57	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	105.57%
24) Chloroform-d	4.65	84	100295	47.52	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.04%
26) 1,2-Dichloroethane-d4	5.31	65	63624	48.91	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.82%
32) Benzene-d6	5.34	84	220648	49.86	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.72%
36) 1,2-Dichloropropane-d6	6.33	67	75449	50.19	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.38%
41) Toluene-d8	7.57	98	192254	47.94	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.88%
43) trans-1,3-Dichloropropene-	7.85	79	30673	45.53	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.06%
47) 2-Hexanone-d5	8.31	63	76270	106.82	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	106.82%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	103511	49.63	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	99.26%
64) 1,2-Dichlorobenzene-d4	11.86	152	64732	51.94	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.88%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU122118\
Data File : VU028906.D
Acq On : 22 Dec 2018 04:20
Operator : MD/SY
Sample : J6542-07
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F9F70

Quant Time: Dec 22 06:07:49 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM122018WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Dec 22 01:55:05 2018
Response via : Initial Calibration

