

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U122120\
 Data File : VU041699.D
 Acq On : 21 Dec 2020 14:26
 Operator : SY/MD
 Sample : L5148-01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GB1T2

Quant Time: Dec 22 03:22:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Dec 22 02:49:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	143086	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	140962	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	63386	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	45102	4.03	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	80.60%
7) Chloroethane-d5	1.92	69	40317	4.66	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.20%
11) 1,1-Dichloroethene-d2	2.58	63	63882	3.10	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.00%
20) 2-Butanone-d5	4.66	46	183223	57.02	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	114.04%
24) Chloroform-d	5.08	84	89395	4.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.20%
26) 1,2-Dichloroethane-d4	5.72	65	56112	5.16	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.20%
32) Benzene-d6	5.74	84	190700	4.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.40%
36) 1,2-Dichloropropane-d6	6.70	67	56884	4.49	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.80%
41) Toluene-d8	7.91	98	150331	4.13	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.60%
43) trans-1,3-Dichloropropene-	8.19	79	26096	4.59	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.80%
46) 2-Hexanone-d5	8.64	63	189332	60.86	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	121.72%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	54362	4.73	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.60%
64) 1,2-Dichlorobenzene-d4	12.19	152	59153	5.25	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.00%

Target Compounds					Ovalue
29) 1,1,1-Trichloroethane	5.33	97	1724	0.106 ug/L	96
31) Carbon tetrachloride	5.55	117	1441	0.103 ug/L	99
47) Tetrachloroethene	8.57	164	14839480	1871.251 ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU122120\
Data File : VU041699.D
Acq On : 21 Dec 2020 14:26
Operator : SY/MD
Sample : L5148-01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GB1T2

Quant Time: Dec 22 03:22:38 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Dec 22 02:49:59 2020
Response via : Initial Calibration

