

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U122120\
 Data File : VU041713.D
 Acq On : 21 Dec 2020 21:41
 Operator : SY/MD
 Sample : VU1221WBL02
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK09

Quant Time: Dec 22 06:07:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Dec 22 05:59:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	139092	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	136659	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	66116	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	58942	5.42	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	108.40%
7) Chloroethane-d5	1.92	69	51947	6.17	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	123.40%
11) 1,1-Dichloroethene-d2	2.58	63	83040	4.15	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	83.00%
20) 2-Butanone-d5	4.66	46	200634	64.23	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	128.46%
24) Chloroform-d	5.08	84	109901	5.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	118.00%
26) 1,2-Dichloroethane-d4	5.72	65	62468	5.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	118.20%
32) Benzene-d6	5.74	84	220306	5.63	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	112.60%
36) 1,2-Dichloropropane-d6	6.70	67	71064	5.79	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	115.80%
41) Toluene-d8	7.90	98	193537	5.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	109.60%
43) trans-1,3-Dichloropropene-	8.19	79	27776	5.04	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	100.80%
46) 2-Hexanone-d5	8.64	63	172257	57.12	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	114.24%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	63601	5.71	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	114.20%
64) 1,2-Dichlorobenzene-d4	12.19	152	68535	5.83	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	116.60%

Target Compounds					Ovalue
16) Methylene chloride	3.06	84	2849	0.253 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU122120\
Data File : VU041713.D
Acq On : 21 Dec 2020 21:41
Operator : SY/MD
Sample : VU1221WBL02
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK09

Quant Time: Dec 22 06:07:23 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Dec 22 05:59:09 2020
Response via : Initial Calibration

