

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122121\
 Data File : VU046537.D
 Acq On : 21 Dec 2021 20:56
 Operator : SY/MD
 Sample : M5149-04
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GBBT4

Manual Integrations
 APPROVED

Reviewed By :Amit Patel 12/24/2021
 Supervised By :Mahesh Dadoda 12/28/2021

Quant Time: Dec 22 06:54:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122121WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Dec 22 06:54:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	78515	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	76875	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	34292	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	31381	5.074	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	101.400%
7) Chloroethane-d5	1.919	69	24900	5.113	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	102.200%
11) 1,1-Dichloroethene-d2	2.575	65	13492	5.174	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	103.400%
20) 2-Butanone-d5	4.658	46	100038	55.766	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	111.540%
24) Chloroform-d	5.070	84	57242	5.051	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.000%
26) 1,2-Dichloroethane-d4	5.706	65	35106	5.419	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.400%
32) Benzene-d6	5.732	84	114635	5.349	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.000%
36) 1,2-Dichloropropane-d6	6.694	67	34427	5.346	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	107.000%
41) Toluene-d8	7.899	98	95970	4.934	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.600%
43) trans-1,3-Dichloroprop...	8.182	79	14812	4.973	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	99.400%
46) 2-Hexanone-d5	8.639	63	95518	54.626	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	109.260%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	33846	5.121	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.400%
66) 1,2-Dichlorobenzene-d4	12.195	152	36458	5.772	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	115.400%
Target Compounds					Qvalue	
13) Acetone	2.687	43	10769m	6.403	ug/L	
30) Cyclohexane	5.395	56	8187	0.779	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122121\
Data File : VU046537.D
Acq On : 21 Dec 2021 20:56
Operator : SY/MD
Sample : M5149-04
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GBBT4

Quant Time: Dec 22 06:54:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122121WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Dec 22 06:54:42 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Amit Patel 12/24/2021
Supervised By :Mahesh Dadoda 12/28/2021

