

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU122122\  
 Data File : VU052473.D  
 Acq On : 21 Dec 2022 12:26  
 Operator : JC/MD  
 Sample : N6156-04  
 Misc : 25.0mL/MSVOA\_U/WATER  
 ALS Vial : 9 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 EXEM0

Quant Time: Dec 22 00:29:56 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\SFAMUTR122022WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Thu Dec 22 00:28:27 2022  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	173966	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	175681	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	71333	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	84584	4.799	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	96.000%	
7) Chloroethane-d5	1.916	69	75214	4.948	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	99.000%	
11) 1,1-Dichloroethene-d2	2.568	65	36209	4.774	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	95.400%	
20) 2-Butanone-d5	4.652	46	263032	48.126	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	96.260%	
24) Chloroform-d	5.060	84	147875	4.973	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	99.400%	
26) 1,2-Dichloroethane-d4	5.697	65	86265	5.001	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	100.000%	
32) Benzene-d6	5.726	84	295562	4.907	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	98.200%	
36) 1,2-Dichloropropane-d6	6.687	67	105207	5.074	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	101.400%	
41) Toluene-d8	7.896	98	227891	4.507	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	90.200%	
43) trans-1,3-Dichloroprop...	8.176	79	33863	4.770	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	95.400%	
46) 2-Hexanone-d5	8.636	63	180950	46.584	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	93.160%	
56) 1,1,2,2-Tetrachloroeth...	10.751	84	84379	4.793	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	95.800%	
66) 1,2-Dichlorobenzene-d4	12.188	152	83520	5.751	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	115.000%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU122122\  
 Data File : VU052473.D  
 Acq On : 21 Dec 2022 12:26  
 Operator : JC/MD  
 Sample : N6156-04  
 Misc : 25.0mL/MSVOA\_U/WATER  
 ALS Vial : 9 Sample Multiplier: 1

**Instrument :**  
 MSVOA\_U  
**ClientSampleId :**  
 EXEM0

Quant Time: Dec 22 00:29:56 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\SFAMUTR122022WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Thu Dec 22 00:28:27 2022  
 Response via : Initial Calibration

