

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122122\
 Data File : VU052473.D
 Acq On : 21 Dec 2022 12:26
 Operator : JC/MD
 Sample : N6156-04
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EXEM0

Quant Time: Dec 22 00:29:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122022WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Dec 22 00:28:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	173966	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	175681	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	71333	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	84584	4.799	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	96.000%
7) Chloroethane-d5	1.916	69	75214	4.948	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.000%
11) 1,1-Dichloroethene-d2	2.568	65	36209	4.774	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	95.400%
20) 2-Butanone-d5	4.652	46	263032	48.126	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.260%
24) Chloroform-d	5.060	84	147875	4.973	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.400%
26) 1,2-Dichloroethane-d4	5.697	65	86265	5.001	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.000%
32) Benzene-d6	5.726	84	295562	4.907	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.200%
36) 1,2-Dichloropropane-d6	6.687	67	105207	5.074	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.400%
41) Toluene-d8	7.896	98	227891	4.507	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.200%
43) trans-1,3-Dichloroprop...	8.176	79	33863	4.770	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.400%
46) 2-Hexanone-d5	8.636	63	180950	46.584	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.160%
56) 1,1,2,2-Tetrachloroeth...	10.751	84	84379	4.793	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.800%
66) 1,2-Dichlorobenzene-d4	12.188	152	83520	5.751	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	115.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122122\
Data File : VU052473.D
Acq On : 21 Dec 2022 12:26
Operator : JC/MD
Sample : N6156-04
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EXEM0

Quant Time: Dec 22 00:29:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122022WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Dec 22 00:28:27 2022
Response via : Initial Calibration

