

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122122\
 Data File : VU052479.D
 Acq On : 21 Dec 2022 16:02
 Operator : JC/MD
 Sample : N6156-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Dec 22 00:31:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122022WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Dec 22 00:28:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	163873	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	167286	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	69276	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	84021	5.061	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 101.200%		
7) Chloroethane-d5	1.912	69	75254	5.256	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 105.200%		
11) 1,1-Dichloroethene-d2	2.568	65	35707	4.998	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 100.000%		
20) 2-Butanone-d5	4.636	46	257260	49.969	ug/L	-0.01
Spiked Amount	50.000	Range 40 - 130	Recovery	= 99.940%		
24) Chloroform-d	5.060	84	146215	5.220	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 104.400%		
26) 1,2-Dichloroethane-d4	5.700	65	86186	5.305	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 106.000%		
32) Benzene-d6	5.726	84	289543	5.048	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 101.000%		
36) 1,2-Dichloropropane-d6	6.687	67	102941	5.214	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 104.200%		
41) Toluene-d8	7.896	98	220720	4.584	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 91.600%		
43) trans-1,3-Dichloroprop...	8.176	79	33139	4.902	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 98.000%		
46) 2-Hexanone-d5	8.632	63	169957	45.950	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 91.900%		
56) 1,1,2,2-Tetrachloroeth...	10.751	84	82499	4.922	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 98.400%		
66) 1,2-Dichlorobenzene-d4	12.189	152	77667	5.507	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 110.200%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122122\
Data File : VU052479.D
Acq On : 21 Dec 2022 16:02
Operator : JC/MD
Sample : N6156-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Dec 22 00:31:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122022WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Dec 22 00:28:27 2022
Response via : Initial Calibration

