

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122123\
 Data File : VU057180.D
 Acq On : 22 Dec 2023 10:08
 Operator : MD/SY
 Sample : 05912-20
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :

Quant Time: Dec 22 12:39:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Dec 22 01:33:10 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	160752	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	157988	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.808	152	68655	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.599	65	76353	4.969	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	99.400%
7) Chloroethane-d5	1.914	69	64490	5.099	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	102.000%
11) 1,1-Dichloroethene-d2	2.564	65	31660	4.756	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	95.200%
20) 2-Butanone-d5	4.625	46	140040	58.050	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	116.100%
24) Chloroform-d	5.056	84	155184	5.230	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.600%
26) 1,2-Dichloroethane-d4	5.695	65	78035	5.518	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	110.400%
32) Benzene-d6	5.721	84	288680	5.327	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.600%
36) 1,2-Dichloropropane-d6	6.682	67	84084	5.453	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	109.000%
41) Toluene-d8	7.891	98	238587	4.823	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.400%
43) trans-1,3-Dichloroprop...	8.174	79	30779	4.795	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.800%
46) 2-Hexanone-d5	8.628	63	114159	52.234	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.460%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	58330	5.316	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.400%
66) 1,2-Dichlorobenzene-d4	12.187	152	80304	5.436	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.800%
Target Compounds						Qvalue
5) Vinyl chloride	1.602	62	11696	0.832	ug/L	86
12) 1,1-Dichloroethene	2.577	96	6961	0.579	ug/L #	1
19) 1,1-Dichloroethane	3.863	63	88956	3.825	ug/L	95
22) cis-1,2-Dichloroethene	4.660	96	96208	7.454	ug/L	99
33) Benzene	5.773	78	8602	0.172	ug/L	100
34) Trichloroethene	6.538	95	24523	1.674	ug/L	93
35) Methylcyclohexane	6.682	83	21130	1.096	ug/L #	19
37) 1,2-Dichloropropane	6.682	63	9612	0.772	ug/L #	88
39) cis-1,3-Dichloropropene	7.560	75	2294	0.122	ug/L #	46
48) 2-Hexanone	8.631	43	14371	3.607	ug/L #	56
51) Chlorobenzene	9.441	112	239741	6.940	ug/L	99
64) 1,3-Dichlorobenzene	11.740	146	3295	0.138	ug/L	94
65) 1,4-Dichlorobenzene	11.830	146	25194	1.046	ug/L	90
67) 1,2-Dichlorobenzene	12.206	146	31050	1.451	ug/L	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122123\
Data File : VU057180.D
Acq On : 22 Dec 2023 10:08
Operator : MD/SY
Sample : 05912-20
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 57 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :


Quant Time: Dec 22 12:39:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Dec 22 01:33:10 2023
Response via : Initial Calibration








#5
 Vinyl chloride
 Concen: 0.832 ug/L
 RT: 1.602 min Scan# 91
 Delta R.T. 0.000 min
 Lab File: VU057180.D
 Acq: 22 Dec 2023 10:08

Instrument :
 MSVOA_U
 ClientSampleId :







Tgt Ion: 62 Resp: 11696
 Ion Ratio Lower Upper
 62 100
 64 41.1 23.3 43.3



#12
 1,1-Dichloroethene
 Concen: 0.579 ug/L
 RT: 2.577 min Scan# 400
 Delta R.T. 0.000 min
 Lab File: VU057180.D
 Acq: 22 Dec 2023 10:08

Tgt Ion: 96 Resp: 6961
 Ion Ratio Lower Upper
 96 100
 61 259.7 119.6 222.0#
 63 861.6 96.2 178.8#









#19
1,1-Dichloroethane
Concen: 3.825 ug/L
RT: 3.863 min Scan# 80
Delta R.T. 0.003 min
Lab File: VU057180.D
Acq: 22 Dec 2023 10:08




Instrument :
MSVOA_U
ClientSampleId :


Tgt Ion: 63 Resp: 88956
Ion Ratio Lower Upper
63 100
65 33.9 21.6 40.2
83 13.2 10.2 19.0



#22
cis-1,2-Dichloroethene
Concen: 7.454 ug/L
RT: 4.660 min Scan# 1048
Delta R.T. 0.000 min
Lab File: VU057180.D
Acq: 22 Dec 2023 10:08

Tgt Ion: 96 Resp: 96208
Ion Ratio Lower Upper
96 100
61 124.2 86.8 161.2
98 64.1 45.5 84.5









#33
Benzene
Concen: 0.172 ug/L
RT: 5.773 min Scan# 11
Delta R.T. 0.007 min
Lab File: VU057180.D
Acq: 22 Dec 2023 10:08

Instrument :
MSVOA_U
ClientSampleId :




Tgt Ion: 78 Resp: 8602




#34
Trichloroethene
Concen: 1.674 ug/L
RT: 6.538 min Scan# 1632
Delta R.T. 0.000 min
Lab File: VU057180.D
Acq: 22 Dec 2023 10:08

Tgt Ion: 95 Resp: 24523

Ion	Ratio	Lower	Upper
95	100		
97	66.5	44.3	82.3
132	96.6	62.6	116.2
130	97.8	62.4	116.0









#35
Methylcyclohexane
Concen: 1.096 ug/L
RT: 6.682 min Scan# 1677
Delta R.T. -0.077 min
Lab File: VU057180.D
Acq: 22 Dec 2023 10:08




Instrument :
MSVOA_U
ClientSampleId :


Tgt Ion: 83 Resp: 21130
Ion Ratio Lower Upper
83 100
55 0.0 61.3 91.9#
98 0.4 36.6 55.0#



#37
1,2-Dichloropropane
Concen: 0.772 ug/L
RT: 6.682 min Scan# 1677
Delta R.T. -0.099 min
Lab File: VU057180.D
Acq: 22 Dec 2023 10:08

Tgt Ion: 63 Resp: 9612
Ion Ratio Lower Upper
63 100
112 0.0 3.1 4.7#









#39
cis-1,3-Dichloropropene
Concen: 0.122 ug/L
RT: 7.560 min Scan# 1963
Delta R.T. -0.042 min
Lab File: VU057180.D
Acq: 22 Dec 2023 10:08




Instrument :
MSVOA_U
ClientSampleId :


Tgt Ion: 75 Resp: 2294
Ion Ratio Lower Upper
75 100
77 61.0 21.8 40.4#



#48
2-Hexanone
Concen: 3.607 ug/L
RT: 8.631 min Scan# 2283
Delta R.T. -0.048 min
Lab File: VU057180.D
Acq: 22 Dec 2023 10:08

Tgt Ion: 43 Resp: 14371
Ion Ratio Lower Upper
43 100
58 23.1 49.8 74.8#
57 38.6 18.5 27.7#
100 0.7 11.1 16.7#









#51
Chlorobenzene
Concen: 6.940 ug/L
RT: 9.441 min Scan# 21
Delta R.T. 0.000 min
Lab File: VU057180.D
Acq: 22 Dec 2023 10:08




Instrument :
MSVOA_U
ClientSampleId :

Tgt Ion:112 Resp: 239741
Ion Ratio Lower Upper
112 100
114 32.6 22.7 42.1
77 63.8 52.2 78.2



#64
1,3-Dichlorobenzene
Concen: 0.138 ug/L
RT: 11.740 min Scan# 3250
Delta R.T. 0.000 min
Lab File: VU057180.D
Acq: 22 Dec 2023 10:08


Tgt Ion:146 Resp: 3295
Ion Ratio Lower Upper
146 100
111 41.0 30.4 56.4
148 68.0 43.4 80.6








#65
1,4-Dichlorobenzene
Concen: 1.046 ug/L
RT: 11.830 min Scan# 31
Delta R.T. 0.000 min
Lab File: VU057180.D
Acq: 22 Dec 2023 10:08

Instrument :
MSVOA_U
ClientSampleId :



Tgt Ion:146 Resp: 25194
Ion Ratio Lower Upper
146 100
111 44.9 28.1 52.3
148 70.0 43.0 79.8



#67
1,2-Dichlorobenzene
Concen: 1.451 ug/L
RT: 12.206 min Scan# 3395
Delta R.T. 0.000 min
Lab File: VU057180.D
Acq: 22 Dec 2023 10:08

Tgt Ion:146 Resp: 31050
Ion Ratio Lower Upper
146 100
111 48.4 37.9 56.9
148 64.9 50.1 75.1

