

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U122220\
 Data File : VU041728.D
 Acq On : 22 Dec 2020 09:45
 Operator : SY/MD
 Sample : VU1222WBL01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK10

Quant Time: Dec 23 04:12:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Dec 23 04:05:15 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	146231	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	141267	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	64543	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	40254	3.52	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	70.40%
7) Chloroethane-d5	1.92	69	45232	5.11	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	102.20%
11) 1,1-Dichloroethene-d2	2.57	63	69930	3.32	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.40%
20) 2-Butanone-d5	4.66	46	183595	55.91	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	111.82%
24) Chloroform-d	5.08	84	91163	4.65	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.00%
26) 1,2-Dichloroethane-d4	5.71	65	52195	4.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.00%
32) Benzene-d6	5.74	84	183734	4.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.80%
36) 1,2-Dichloropropane-d6	6.70	67	59832	4.72	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.40%
41) Toluene-d8	7.90	98	152997	4.19	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.80%
43) trans-1,3-Dichloropropene-	8.18	79	24981	4.38	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.60%
46) 2-Hexanone-d5	8.64	63	161572	51.83	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.66%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	59890	5.20	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.00%
64) 1,2-Dichlorobenzene-d4	12.19	152	58645	5.11	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.20%

Target Compounds					Ovalue
16) Methylene chloride	3.06	84	2040	0.172 ug/L	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU122220\
Data File : VU041728.D
Acq On : 22 Dec 2020 09:45
Operator : SY/MD
Sample : VU1222WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK10

Quant Time: Dec 23 04:12:32 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Dec 23 04:05:15 2020
Response via : Initial Calibration

