

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U122220\
 Data File : VU041731.D
 Acq On : 22 Dec 2020 11:05
 Operator : SY/MD
 Sample : L5148-04
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GB1T3

Quant Time: Dec 23 04:23:50 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Dec 23 04:05:15 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	137459	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	133694	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	58837	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	41123	3.83	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	76.60%
7) Chloroethane-d5	1.92	69	44807	5.39	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	107.80%
11) 1,1-Dichloroethene-d2	2.58	63	71218	3.60	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.00%
20) 2-Butanone-d5	4.65	46	179494	58.15	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	116.30%
24) Chloroform-d	5.08	84	96469	5.24	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.80%
26) 1,2-Dichloroethane-d4	5.72	65	54872	5.25	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.00%
32) Benzene-d6	5.74	84	188250	4.92	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.40%
36) 1,2-Dichloropropane-d6	6.70	67	60492	5.04	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.80%
41) Toluene-d8	7.90	98	157664	4.56	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.20%
43) trans-1,3-Dichloropropene-	8.19	79	23909	4.43	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.60%
46) 2-Hexanone-d5	8.64	63	161131	54.61	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	109.22%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	58892	5.40	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	108.00%
64) 1,2-Dichlorobenzene-d4	12.19	152	60391	5.77	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	115.40%

Target Compounds					Ovalue
31) Carbon tetrachloride	5.54	117	6620	0.499 ug/L	94
47) Tetrachloroethene	8.56	164	52024	6.917 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU122220\
Data File : VU041731.D
Acq On : 22 Dec 2020 11:05
Operator : SY/MD
Sample : L5148-04
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GB1T3

Quant Time: Dec 23 04:23:50 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Dec 23 04:05:15 2020
Response via : Initial Calibration

