

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122222\
 Data File : VU052490.D
 Acq On : 22 Dec 2022 14:42
 Operator : JC/MD
 Sample : N6170-11
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GBQB8

Quant Time: Dec 23 00:32:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122022WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Dec 22 00:28:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	203558	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	198541	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	98163	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	79385	3.849	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	77.000%
7) Chloroethane-d5	1.916	69	75466	4.243	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	84.800%
11) 1,1-Dichloroethene-d2	2.568	65	34604	3.899	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.000%
20) 2-Butanone-d5	4.649	46	301762	47.186	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	94.380%
24) Chloroform-d	5.060	84	151782	4.362	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.200%
26) 1,2-Dichloroethane-d4	5.700	65	90588	4.489	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.800%
32) Benzene-d6	5.723	84	318848	4.684	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.600%
36) 1,2-Dichloropropane-d6	6.687	67	111727	4.768	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.400%
41) Toluene-d8	7.896	98	260011	4.550	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.000%
43) trans-1,3-Dichloroprop...	8.176	79	37988	4.735	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	94.600%
46) 2-Hexanone-d5	8.632	63	214302	48.818	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.640%
56) 1,1,2,2-Tetrachloroeth...	10.751	84	90295	4.539	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.800%
66) 1,2-Dichlorobenzene-d4	12.189	152	100841	5.046	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122222\
Data File : VU052490.D
Acq On : 22 Dec 2022 14:42
Operator : JC/MD
Sample : N6170-11
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GBQB8

Quant Time: Dec 23 00:32:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122022WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Dec 22 00:28:27 2022
Response via : Initial Calibration

