

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122223\
 Data File : VU057213.D
 Acq On : 23 Dec 2023 01:25
 Operator : MD/SY
 Sample : 05927-07
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BGWA7

Quant Time: Dec 23 04:42:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Dec 23 00:27:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	180584	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	177914	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	79112	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.599	65	69942	4.052	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	81.000%
7) Chloroethane-d5	1.914	69	62987	4.433	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	88.600%
11) 1,1-Dichloroethene-d2	2.563	65	31102	4.159	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	83.200%
20) 2-Butanone-d5	4.624	46	141536	52.227	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.460%
24) Chloroform-d	5.055	84	150892	4.527	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.600%
26) 1,2-Dichloroethane-d4	5.695	65	71683	4.512	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.200%
32) Benzene-d6	5.721	84	270756	4.437	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.800%
36) 1,2-Dichloropropane-d6	6.685	67	84838	4.886	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.800%
41) Toluene-d8	7.894	98	222486	3.993	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.800%
43) trans-1,3-Dichloroprop...	8.174	79	29648	4.101	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.000%
46) 2-Hexanone-d5	8.627	63	117907	47.907	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.820%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	62841	5.086	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.800%
66) 1,2-Dichlorobenzene-d4	12.187	152	82317	4.835	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.800%
Target Compounds						Qvalue
13) Acetone	2.641	43	18236	11.192	ug/L	99
42) Toluene	7.968	91	37356	0.625	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122223\
Data File : VU057213.D
Acq On : 23 Dec 2023 01:25
Operator : MD/SY
Sample : 05927-07
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BGWA7

Quant Time: Dec 23 04:42:34 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Dec 23 00:27:07 2023
Response via : Initial Calibration

