

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122223\
 Data File : VU057229.D
 Acq On : 23 Dec 2023 07:53
 Operator : MD/SY
 Sample : 05979-13
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BGR43

Quant Time: Dec 25 00:16:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Dec 23 00:27:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	166465	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	167627	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	70077	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	68097	4.280	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	85.600%
7) Chloroethane-d5	1.910	69	59139	4.515	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	90.400%
11) 1,1-Dichloroethene-d2	2.563	65	29069	4.217	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	84.400%
20) 2-Butanone-d5	4.618	46	147827	59.175	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	118.360%
24) Chloroform-d	5.055	84	148836	4.844	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.800%
26) 1,2-Dichloroethane-d4	5.695	65	72791	4.971	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.400%
32) Benzene-d6	5.721	84	265499	4.618	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.400%
36) 1,2-Dichloropropane-d6	6.682	67	86014	5.257	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.200%
41) Toluene-d8	7.894	98	214276	4.082	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.600%
43) trans-1,3-Dichloroprop...	8.174	79	29213	4.289	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.800%
46) 2-Hexanone-d5	8.627	63	119100	51.361	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.720%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	61795	5.308	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.200%
66) 1,2-Dichlorobenzene-d4	12.187	152	74909	4.968	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.400%
Target Compounds					Qvalue	
13) Acetone	2.637	43	8093	5.388	ug/L	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122223\
Data File : VU057229.D
Acq On : 23 Dec 2023 07:53
Operator : MD/SY
Sample : 05979-13
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BGR43

Quant Time: Dec 25 00:16:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Dec 23 00:27:07 2023
Response via : Initial Calibration





#13

Acetone

Concen: 5.388 ug/L

RT: 2.637 min Scan# 41

Delta R.T. -0.000 min

Lab File: VU057229.D


Acq: 23 Dec 2023 07:53

Instrument :

MSVOA_U

ClientSampleId :

BGR43



Tgt Ion: 43 Resp: 8093

Ion Ratio Lower Upper

43 100

58 36.6 0.0 70.8

