

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122321\
 Data File : VU046571.D
 Acq On : 23 Dec 2021 17:25
 Operator : SY/MD
 Sample : M5170-11
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 H0A07

Quant Time: Dec 24 04:56:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122121WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Dec 24 04:54:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	81681	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	81422	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.816	152	34983	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	21226	3.299	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	66.000%	
7) Chloroethane-d5	1.919	69	19180	3.786	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	75.800%	
11) 1,1-Dichloroethene-d2	2.572	65	9144	3.370	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	67.400%	
20) 2-Butanone-d5	4.658	46	86935	46.584	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	93.160%	
24) Chloroform-d	5.070	84	47500	4.029	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	80.600%	
26) 1,2-Dichloroethane-d4	5.706	65	28927	4.292	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	85.800%	
32) Benzene-d6	5.732	84	87296	3.846	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	77.000%	
36) 1,2-Dichloropropane-d6	6.694	67	28511	4.180	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	83.600%	
41) Toluene-d8	7.899	98	70245	3.410	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	68.200%#	
43) trans-1,3-Dichloroprop...	8.182	79	11691	3.706	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	74.200%	
46) 2-Hexanone-d5	8.639	63	88983	48.047	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	96.100%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	31043	4.435	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	88.600%	
66) 1,2-Dichlorobenzene-d4	12.195	152	30126	4.675	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	93.600%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122321\
 Data File : VU046571.D
 Acq On : 23 Dec 2021 17:25
 Operator : SY/MD
 Sample : M5170-11
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
ClientSampleId :
 H0A07

Quant Time: Dec 24 04:56:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122121WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Dec 24 04:54:59 2021
 Response via : Initial Calibration

