

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122324\
 Data File : VU062561.D
 Acq On : 23 Dec 2024 12:08
 Operator : MD/SY
 Sample : VU1223WBL01
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK125

Quant Time: Dec 23 14:53:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Dec 20 23:41:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	101638	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	92715	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	41854	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.599	65	26359	3.845	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	76.800%
7) Chloroethane-d5	1.914	69	22142	4.541	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	90.800%
11) 1,1-Dichloroethene-d2	2.563	65	14219	4.360	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	87.200%
20) 2-Butanone-d5	4.644	46	59159	45.027	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	90.060%
24) Chloroform-d	5.052	84	70331	4.740	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.800%
26) 1,2-Dichloroethane-d4	5.692	65	38869	5.327	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.600%
32) Benzene-d6	5.718	84	124336	4.656	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.200%
36) 1,2-Dichloropropane-d6	6.679	67	35292	4.619	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.400%
41) Toluene-d8	7.888	98	119558	4.740	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.800%
43) trans-1,3-Dichloroprop...	8.174	79	15051	4.475	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.600%
46) 2-Hexanone-d5	8.631	63	45950	45.288	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.580%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	27594	4.641	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.800%
66) 1,2-Dichlorobenzene-d4	12.184	152	41839	5.425	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.400%
Target Compounds					Qvalue	
34) Trichloroethene	6.534	95	11686	1.436	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122324\
Data File : VU062561.D
Acq On : 23 Dec 2024 12:08
Operator : MD/SY
Sample : VU1223WBL01
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK125

Quant Time: Dec 23 14:53:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Dec 20 23:41:56 2024
Response via : Initial Calibration

