

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU122418\
 Data File : VU028947.D
 Acq On : 24 Dec 2018 15:06
 Operator : MD/SY
 Sample : J6542-10
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F9F78

Quant Time: Dec 26 09:42:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM122018WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Dec 25 00:02:36 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	258786	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	246076	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	119149	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	77804	37.50	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	75.00%
7) Chloroethane-d5	1.67	69	67275	41.98	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	83.96%
11) 1,1-Dichloroethene-d2	2.27	63	111866	32.06	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	64.12%
21) 2-Butanone-d5	4.18	46	131943	90.77	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	90.77%
24) Chloroform-d	4.65	84	149209	45.74	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.48%
26) 1,2-Dichloroethane-d4	5.31	65	94736	47.11	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.22%
32) Benzene-d6	5.34	84	332730	47.86	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.72%
36) 1,2-Dichloropropane-d6	6.33	67	110081	46.61	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.22%
41) Toluene-d8	7.56	98	311484	49.44	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.88%
43) trans-1,3-Dichloropropene-	7.85	79	50631	47.84	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.68%
47) 2-Hexanone-d5	8.31	63	114872	102.41	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	102.41%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	153537	46.86	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.72%
64) 1,2-Dichlorobenzene-d4	11.86	152	123248	51.14	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.28%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU122418\
Data File : VU028947.D
Acq On : 24 Dec 2018 15:06
Operator : MD/SY
Sample : J6542-10
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F9F78

Quant Time: Dec 26 09:42:30 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM122018WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Dec 25 00:02:36 2018
Response via : Initial Calibration

