

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U122520\
 Data File : VU041814.D
 Acq On : 24 Dec 2020 18:17
 Operator : SY/MD
 Sample : L5185-08
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Dec 25 05:03:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Dec 25 03:47:04 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	126696	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	123875	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	54217	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	40908	4.13	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.60%
7) Chloroethane-d5	1.92	69	40731	5.31	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	106.20%
11) 1,1-Dichloroethene-d2	2.58	63	61527	3.37	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.40%
20) 2-Butanone-d5	4.66	46	154325	54.24	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.48%
24) Chloroform-d	5.08	84	87511	5.16	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.20%
26) 1,2-Dichloroethane-d4	5.72	65	48855	5.08	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.60%
32) Benzene-d6	5.74	84	162226	4.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.40%
36) 1,2-Dichloropropane-d6	6.70	67	52355	4.71	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.20%
41) Toluene-d8	7.91	98	128541	4.01	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.20%
43) trans-1,3-Dichloropropene-	8.19	79	20295	4.06	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.20%
46) 2-Hexanone-d5	8.64	63	127312	46.57	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.14%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	49625	4.91	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.20%
64) 1,2-Dichlorobenzene-d4	12.19	152	50371	5.22	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.40%

Target Compounds					Ovalue
16) Methylene chloride	3.06	84	2157	0.210 ug/L	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU122520\
Data File : VU041814.D
Acq On : 24 Dec 2020 18:17
Operator : SY/MD
Sample : L5185-08
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Dec 25 05:03:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Dec 25 03:47:04 2020
Response via : Initial Calibration

