

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122623\
 Data File : VU057278.D
 Acq On : 27 Dec 2023 10:37
 Operator : MD/SY
 Sample : 06002-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AF9

Quant Time: Dec 28 00:56:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Dec 27 01:05:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	160788	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	171633	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	74654	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	90642	5.898	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	118.000%
7) Chloroethane-d5	1.911	69	70299	5.557	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	111.200%
11) 1,1-Dichloroethene-d2	2.563	65	34963	5.251	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	105.000%
20) 2-Butanone-d5	4.624	46	150343	62.307	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	124.620%
24) Chloroform-d	5.055	84	147208	4.960	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.200%
26) 1,2-Dichloroethane-d4	5.695	65	73054	5.165	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.200%
32) Benzene-d6	5.721	84	279417	4.747	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.000%
36) 1,2-Dichloropropane-d6	6.682	67	87656	5.233	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.600%
41) Toluene-d8	7.891	98	226080	4.206	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.200%
43) trans-1,3-Dichloroprop...	8.177	79	31108	4.461	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.200%
46) 2-Hexanone-d5	8.627	63	113211	47.682	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.360%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	61748	5.180	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.600%
66) 1,2-Dichlorobenzene-d4	12.187	152	78006	4.856	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.200%
Target Compounds						
					Qvalue	
3) Chloromethane	1.518	50	7703	0.585	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122623\
Data File : VU057278.D
Acq On : 27 Dec 2023 10:37
Operator : MD/SY
Sample : 06002-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AF9

Quant Time: Dec 28 00:56:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Dec 27 01:05:12 2023
Response via : Initial Calibration

