

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122624\
 Data File : VU062578.D
 Acq On : 26 Dec 2024 17:31
 Operator : MD/SY
 Sample : P5353-01
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E2927

Quant Time: Dec 27 00:26:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Dec 27 00:22:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	94113	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	89967	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	45642	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	22214	3.499	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	70.000%
7) Chloroethane-d5	1.907	69	19175	4.247	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	85.000%
11) 1,1-Dichloroethene-d2	2.560	65	11909	3.943	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.800%
20) 2-Butanone-d5	4.624	46	53019	43.580	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	87.160%
24) Chloroform-d	5.052	84	56830	4.136	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.800%
26) 1,2-Dichloroethane-d4	5.692	65	33108	4.901	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.000%
32) Benzene-d6	5.717	84	102176	3.943	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.800%
36) 1,2-Dichloropropane-d6	6.679	67	28618	3.860	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	77.200%
41) Toluene-d8	7.888	98	94227	3.849	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	77.000%
43) trans-1,3-Dichloroprop...	8.174	79	13254	4.061	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.200%
46) 2-Hexanone-d5	8.627	63	48618	49.382	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.760%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	24606	4.265	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	85.400%
66) 1,2-Dichlorobenzene-d4	12.183	152	36603	4.352	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	87.000%
Target Compounds					Qvalue	
13) Acetone	2.640	43	11697	14.121	ug/L	97
25) Chloroform	5.074	83	2436	0.181	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122624\
Data File : VU062578.D
Acq On : 26 Dec 2024 17:31
Operator : MD/SY
Sample : P5353-01
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E2927

Quant Time: Dec 27 00:26:40 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Dec 27 00:22:56 2024
Response via : Initial Calibration

