

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122624\
 Data File : VU062581.D
 Acq On : 26 Dec 2024 18:44
 Operator : MD/SY
 Sample : P5353-04
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E2930

Quant Time: Dec 27 00:27:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Dec 27 00:22:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	92810	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	85454	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	43870	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	21345	3.410	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	68.200%
7) Chloroethane-d5	1.907	69	18971	4.260	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	85.200%
11) 1,1-Dichloroethene-d2	2.560	65	11794	3.960	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	79.200%
20) 2-Butanone-d5	4.624	46	62251	51.887	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.780%
24) Chloroform-d	5.052	84	62352	4.602	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.000%
26) 1,2-Dichloroethane-d4	5.692	65	35856	5.382	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.600%
32) Benzene-d6	5.718	84	106263	4.318	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.400%
36) 1,2-Dichloropropane-d6	6.679	67	30808	4.375	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	87.400%
41) Toluene-d8	7.888	98	101600	4.370	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.400%
43) trans-1,3-Dichloroprop...	8.174	79	14401	4.646	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.000%
46) 2-Hexanone-d5	8.624	63	56673	60.603	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	121.200%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	27865	5.085	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.800%
66) 1,2-Dichlorobenzene-d4	12.184	152	41995	5.195	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.800%
Target Compounds					Qvalue	
13) Acetone	2.640	43	14618	17.895	ug/L	93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122624\
Data File : VU062581.D
Acq On : 26 Dec 2024 18:44
Operator : MD/SY
Sample : P5353-04
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E2930

Quant Time: Dec 27 00:27:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Dec 27 00:22:56 2024
Response via : Initial Calibration





#13
 Acetone
 Concen: 17.895 ug/L
 RT: 2.640 min Scan# 42
 Delta R.T. 0.000 min
 Lab File: VU062581.D
 Acq: 26 Dec 2024 18:44

Instrument :
 MSVOA_U
 ClientSampleId :
 E2930

Tgt Ion: 43 Resp: 14618
 Ion Ratio Lower Upper
 43 100
 58 28.1 0.0 64.0

