

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U122718\
 Data File : VU028955.D
 Acq On : 27 Dec 2018 09:30
 Operator : MD/SY
 Sample : VU1227MBL01
 Misc : 5.00g/5mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK46

Quant Time: Dec 28 00:57:27 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM122018WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 28 00:46:50 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	248372	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	229894	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	103432	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	88998	44.70	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	89.40%
7) Chloroethane-d5	1.67	69	76599	49.81	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.62%
11) 1,1-Dichloroethene-d2	2.27	63	130483	38.97	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	77.94%
21) 2-Butanone-d5	4.17	46	136020	97.50	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.50%
24) Chloroform-d	4.65	84	159577	50.96	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.92%
26) 1,2-Dichloroethane-d4	5.30	65	97978	50.76	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.52%
32) Benzene-d6	5.34	84	344705	53.07	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.14%
36) 1,2-Dichloropropane-d6	6.33	67	115241	52.23	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	104.46%
41) Toluene-d8	7.56	98	316732	53.81	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	107.62%
43) trans-1,3-Dichloropropene-	7.85	79	52849	53.46	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	106.92%
47) 2-Hexanone-d5	8.31	63	114596	109.35	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	109.35%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	160930	52.57	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	105.14%
64) 1,2-Dichlorobenzene-d4	11.86	152	118386	56.59	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	113.18%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU122718\
Data File : VU028955.D
Acq On : 27 Dec 2018 09:30
Operator : MD/SY
Sample : VU1227MBL01
Misc : 5.00µ/5mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK46

Quant Time: Dec 28 00:57:27 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM122018WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 28 00:46:50 2018
Response via : Initial Calibration

