

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122722\
 Data File : VU052571.D
 Acq On : 27 Dec 2022 17:48
 Operator : JC/MD
 Sample : N6170-07DL 5X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GBQB4DL

Quant Time: Dec 28 01:23:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122022WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Dec 28 01:18:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	182309	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	176971	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.806	152	78327	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	56884	3.080	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	61.600%
7) Chloroethane-d5	1.916	69	58765	3.689	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	73.800%
11) 1,1-Dichloroethene-d2	2.568	65	27576	3.470	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.400%
20) 2-Butanone-d5	4.639	46	284684	49.704	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.400%
24) Chloroform-d	5.060	84	130099	4.175	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.400%
26) 1,2-Dichloroethane-d4	5.700	65	76914	4.255	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.200%
32) Benzene-d6	5.726	84	254118	4.188	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.800%
36) 1,2-Dichloropropane-d6	6.687	67	92385	4.423	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.400%
41) Toluene-d8	7.893	98	196182	3.852	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	77.000%
43) trans-1,3-Dichloroprop...	8.176	79	30692	4.291	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.800%
46) 2-Hexanone-d5	8.629	63	193334	49.410	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.820%
56) 1,1,2,2-Tetrachloroeth...	10.751	84	79093	4.460	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.200%
66) 1,2-Dichlorobenzene-d4	12.188	152	77030	4.830	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.600%
Target Compounds						
					Qvalue	
5) Vinyl chloride	1.604	62	3243	0.153	ug/L	87
16) Methylene chloride	3.047	84	22304	1.156	ug/L	98
18) trans-1,2-Dichloroethene	3.353	96	1909	0.136	ug/L	84
22) cis-1,2-Dichloroethene	4.661	96	59947	3.882	ug/L	98
47) Tetrachloroethene	8.549	164	1435	0.138	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122722\
Data File : VU052571.D
Acq On : 27 Dec 2022 17:48
Operator : JC/MD
Sample : N6170-07DL 5X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GBQB4DL

Quant Time: Dec 28 01:23:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122022WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Dec 28 01:18:16 2022
Response via : Initial Calibration

